Spread Spectrum Magnetic Resonance Imaging

Gilles Puy*, Jose P. Marques, Rolf Gruetter, Jean-Philippe Thiran, Senior Member, IEEE, Dimitri Van De Ville, Member, IEEE, Pierre Vandergheynst, and Yves Wiaux, Member, IEEE

Abstract—We propose a novel compressed sensing technique to accelerate the magnetic resonance imaging (MRI) acquisition process. The method, coined spread spectrum MRI or simply s_2 MRI, consists of premodulating the signal of interest by a linear chirp before random k-space under-sampling, and then reconstructing the signal with nonlinear algorithms that promote sparsity. The effectiveness of the procedure is theoretically underpinned by the optimization of the coherence between the sparsity and sensing bases. The proposed technique is thoroughly studied by means of numerical simulations, as well as phantom and *in vivo* experiments on a 7T scanner. Our results suggest that s_2 MRI performs better than state-of-the-art variable density k-space under-sampling approaches.

Index Terms—Compressed sensing, magnetic resonance imaging (MRI), spread spectrum.

I. INTRODUCTION

I N magnetic resonance imaging (MRI), the signal of interest ρ represents the magnetization induced by resonance in the imaged tissues. In the standard setting, data acquired in MRI provide complete Fourier, or *k*-space, measurements of this

Manuscript received June 10, 2011; revised September 06, 2011, October 11, 2011; accepted October 11, 2011 Date of publication October 26, 2011; date of current version March 02, 2012. This work is supported in part by the Center for Biomedical Imaging (CIBM) of the Geneva and Lausanne Universities, EPFL, and the Leenaards and Louis-Jeantet foundations, in part by the Swiss National Science Foundation (SNSF) under Grant PP00P2-123438, and also by the EPFL-Merck Serono Alliance award. *Asterisk indicates corresponding author.*

*G. Puy is with the Institute of Electrical Engineering, Ecole Polytechnique Fédérale de Lausanne (EPFL), CH-1015 Lausanne, Switzerland, and also with the Institute of the Physics of Biological Systems, Ecole Polytechnique Fédérale de Lausanne (EPFL), CH-1015 Lausanne, Switzerland (e-mail: gilles.puy@epfl.ch).

J. P. Marques is with the Institute of the Physics of Biological Systems, Ecole Polytechnique Fédérale de Lausanne (EPFL), CH-1015 Lausanne, Switzerland, and also with the Department of Radiology, University of Lausanne (UNIL), CH-1015 Lausanne, Switzerland.

R. Gruetter is with the Institute of the Physics of Biological Systems, Ecole Polytechnique Fédérale de Lausanne (EPFL), CH-1015 Lausanne, Switzerland, and with the Department of Radiology, University of Lausanne (UNIL), CH-1015 Lausanne, Switzerland, and also with the Department of Radiology and Medical Informatics, University of Geneva (UniGE), CH-1211 Geneva, Switzerland.

J.-P. Thiran and P. Vandergheynst are with the Institute of Electrical Engineering, Ecole Polytechnique Fédérale de Lausanne (EPFL), CH-1015 Lausanne, Switzerland.

D. Van De Ville is with the Institute of Bioengineering, Ecole Polytechnique Fédérale de Lausanne (EPFL), CH-1015 Lausanne, Switzerland, and also with the Department of Radiology and Medical Informatics, University of Geneva (UniGE), CH-1211 Geneva, Switzerland.

Y. Wiaux is with the Institute of Electrical Engineering, Ecole Polytechnique Fédérale de Lausanne (EPFL), CH-1015 Lausanne, Switzerland, and with the Institute of Bioengineering, Ecole Polytechnique Fédérale de Lausanne (EPFL), CH-1015 Lausanne, Switzerland, and also with the Department of Radiology and Medical Informatics, University of Geneva (UniGE), CH-1211 Geneva, Switzerland (e-mail: yves.wiaux@epfl.ch).

Digital Object Identifier 10.1109/TMI.2011.2173698

signal ρ . Accelerating the acquisition process, or equivalently increasing the achievable resolution for a fixed acquisition time, is of major interest for MRI applications. Recent approaches based on compressed sensing seek to reconstruct the signal from incomplete *k*-space information, hence defining an ill-posed inverse problem. The ill-posed inverse problem is regularized by the introduction of sparsity priors, acknowledging the fact that many MRI signals are sparse in well-chosen bases; i.e., that their expansion contains only a small number of nonzero coefficients. Such compressed sensing techniques have been developed for static and dynamic imaging [1]–[5], parallel MRI [6]–[8], MR spectroscopic imaging [9]–[11], and many other applications. Several algorithms have also been proposed to reconstruct MRI signals from under-sampled *k*-space measurements (see, e.g., [12]–[14]).

In the framework of compressed sensing, signals are usually measured through random matrices to ensure that any sparse signal can be recovered with overwhelming probability. The common approach in MRI consists in also exploiting the fact that the energy of MRI signals is usually concentrated at low spatial frequencies. Therefore a variable density k-space random sampling where the under-sampling ratio increases at high frequencies is usually used [1]. This method, heuristic in nature, was shown to be very effective in enhancing the signal reconstruction quality when random distributions optimizing the associated point spread function are used. Other approaches that optimize the acquisition procedure have also been proposed: k-space sampling optimization by Bayesian inference [15]; random k-space convolution with Toeplitz matrices [16]–[18]; encoding by projection onto random waveforms with Gaussian distributions [19].

In the present work, we propose the use of a spread spectrum technique to accelerate single coil MRI acquisition in the framework of compressed sensing. We study this method, coined spread spectrum MRI or simply s₂MRI, theoretically, numerically via simulations, and empirically via real acquisitions. The essence of our strategy consists of premodulating the image by a linear chirp, which results from the application of quadratic phase profiles, and then performing random k-space under-sampling. Images are then reconstructed with nonlinear algorithms promoting signal sparsity. The enhancement of the signal reconstruction quality is linked to a decrease of coherence of the measurement system [20], [21]. In MRI, this type of modulation is known as phase scrambling. It can be obtained by using dedicated coils or by modifying radio-frequency (RF) pulses. It has been used for various purposes, such as improving dynamic range [22], [23], or reducing aliasing artifacts [24], [25], but never in a compressed sensing perspective.

Let us acknowledge that this spread spectrum technique was initially introduced by some of the authors for compressive sampling of pulse trains in [26]. Its transfer to a setting encompassing analog signals and modulations was studied in the context of radio interferometry [21], [27], [28]. The effectiveness of the method for MRI was briefly discussed in [29]–[31].

The paper is organized as follows. In Section II, we explain the principle of the spread spectrum technique in a simplified analog setting. In Section III, we formulate the inverse problem for image reconstruction from under-sampled k-space measurements in the presence of chirp modulation and compare the s₂MRI technique to the variable density sampling method on the basis of numerical simulations of MR acquisitions of a brain image. In Section IV, we describe our implementation of the chirp modulation (on a 7T scanner) and show the effectiveness of s₂MRI using numerical simulations in this precise setting and real acquisitions of phantom and *in vivo* data. Finally, we conclude and discuss potential evolutions of the technique in Section V.

II. SPREAD SPECTRUM PRINCIPLE

A. Compressed Sensing

The essence of the recent theory of compressed sensing is the merging of data acquisition and compression [32]–[38]. Beyond MRI, it is well known that a large variety of natural signals are sparse or compressible in multiscale bases, such as wavelet bases. In the compressed sensing theory, signals are usually expressed as N-dimensional vectors: $\rho \in \mathbb{C}^N$. By definition, a signal is sparse in some orthonormal basis $\Psi \in \mathbb{C}^{N \times N}$, called a sparsity basis, if its expansion contains only a small number $K \ll N$ of nonzero coefficients. More generally it is compressible if its expansion contains only a small number of significant coefficients. The decomposition of ρ in Ψ is denoted $\alpha \in \mathbb{C}^N$, and it satisfies

$$\boldsymbol{\rho} = \boldsymbol{\Psi} \boldsymbol{\alpha}. \tag{1}$$

The theory of compressed sensing demonstrates that a small number $M \ll N$ of linear and nonadaptative measurements $\boldsymbol{\nu} \in \mathbb{C}^M$ suffices for an accurate and stable reconstruction of the signal $\boldsymbol{\rho}$. These measurements may, for example, be obtained by projection onto M randomly selected basis vectors of an orthonormal basis $\boldsymbol{\Phi} \in \mathbb{C}^{N \times N}$, called a sensing basis. This selection process can be modeled by a multiplication with a rectangular binary matrix $\mathbf{M} \in \mathbb{R}^{M \times N}$ that contains only one nonzero value on each line, at the index of the basis vector to be selected. To model a nonperfect sensing process, these measurements are assumed to be contaminated by independent and identically distributed noise $\boldsymbol{n} \in \mathbb{C}^M$. The measurement model thus satisfies

$$\boldsymbol{\nu} = \boldsymbol{\Theta}\boldsymbol{\alpha} + \boldsymbol{n}, \text{with } \boldsymbol{\Theta} = \mathsf{M}\boldsymbol{\Phi}^{\star}\boldsymbol{\Psi} \in \mathbb{R}^{M \times N}$$
(2)

where the matrix Θ identifies the measurement matrix as seen from the sparsity basis (\cdot^* denotes the conjugate transpose operation).

To reconstruct the signal ρ from the measurements ν , the compressed sensing framework proposes, among other approaches, to solve the Basis Pursuit (BP) minimization problem [32]–[39]. This problem regularizes the originally ill-posed inverse problem related to (2) with an explicit sparsity or

compressibility prior on the signal. In the presence of noise, the BP problem is the minimization of the ℓ_1 norm¹ of α under a constraint on the ℓ_2 norm² of the residual noise $\bar{n} = \nu - \Theta \bar{\alpha}$

$$\boldsymbol{\alpha}^* = \arg\min_{\bar{\boldsymbol{\alpha}} \in \mathbb{C}^N} ||\bar{\boldsymbol{\alpha}}||_1 \text{ subject to } ||\boldsymbol{\nu} - \Theta \bar{\boldsymbol{\alpha}}||_2 \le \epsilon.$$
(3)

The corresponding reconstructed signal is $\rho^* = \Psi \alpha^*$.

In this setting, the compressed sensing theory shows that if the number of measurements satisfies

$$M \ge DNK\mu^2(\Phi^\star, \Psi) \ln^4 N \tag{4}$$

for a universal constant D, then α is recovered accurately by solving the BP problem (3) [39]. In relation (4), $\mu(\Phi^*, \Psi)$ is the mutual coherence between the sparsity and sensing bases. It is defined as the maximum projection, in absolute value, between the sparsity basis vectors φ_i , $1 \le i \le N$, and sensing bases vectors $\psi_{i'}$, $1 \le i' \le N$ [36], [39]

$$N^{-1/2} \le \mu(\Phi^{\star}, \Psi) = \max_{1 \le i, i' \le N} |\boldsymbol{\varphi}_i \cdot \boldsymbol{\psi}_{i'}| \le 1.$$
 (5)

The mutual coherence $\mu(\Phi^*, \Psi)$ plays a crucial role in relation (4). Indeed, the number of measurements M^* needed to reconstruct K-sparse signals increases quadratically with its value: $M^* \propto \mu^2(\Phi^*, \Psi)$. In the worst case where $\mu(\Phi^{\star}, \Psi) = 1, M^{\star}$ is of the order of the signal dimension N and under-sampling is impossible. However, when the mutual coherence is at its minimum, $N^{-1/2}$, M^* is reduced to the order of the sparsity level K. This result can intuitively be explained with a simple consideration on the spread of the energy of the sparsity basis vectors in the measurement domain Φ . In an incoherent orthonormal system, the absolute value of the scalar product between the sensing basis vector φ_i and the sparsity basis vector $\boldsymbol{\psi}_{i'}$ is small for all pairs of indices $1 \leq i, i' \leq N$. As $\| \Phi^{\star} \psi_{i'} \|_{2} = \| \psi_{i'} \|_{2}$, the energy of the sparsity basis vector $\boldsymbol{\psi}_{i'}$ spreads equally over the sensing basis vectors φ_i . Consequently, whatever index i is selected to perform a measurement, one always gets information concerning all the sparsity basis vectors describing the original signal. The number of measurements needed for accurate recovery thus decreases.

B. Spread Spectrum and Coherence Reduction

The above considerations suggest that to reduce to the number of measurements needed for accurate recovery of MRI signals, one can try to modify the MRI acquisition procedure in such a way that the energy of the sparsity basis vectors spreads all over the k-space. Following this idea, the s₂MRI technique introduces a linear chirp modulation of the signal of interest before random selection of k-space coefficients (see Fig. 1). Indeed, this modulation conserves the energy of the input signal and corresponds to a convolution that generically spreads the spectrum.

To study theoretically the proposed acquisition scheme, we consider a simplified analog setting where the 1-D signal of interest is denoted by a complex-valued function $\rho(x)$ of the position $x \in \mathbb{R}$. This signal is limited on a finite field-of-view L. For

 $\begin{aligned} &||\boldsymbol{\alpha}||_{1} = \sum_{i=1}^{N} |\alpha_{i}| \text{ where } \alpha_{i} \text{ is the } i\text{th entry of the vector } \boldsymbol{\alpha}. \\ &2||\boldsymbol{n}||_{2} = (\sum_{i=1}^{M} |n_{i}|^{2})^{1/2} \text{ where } n_{i} \text{ is the } i\text{th entry of the vector } \boldsymbol{n}. \end{aligned}$



Fig. 1. Spread spectrum principle. A signal ρ (top left panel) is modulated by a linear chirp (top right panel). In **k**-space, the modulation amounts to the convolution of the Fourier transform of the signal (middle left panel) with that of the chirp (middle right panel). The spectrum of the resulting signal (bottom panel) spreads out in **k**-space.

the sake of simplicity of the following theoretical result only, we assume that the energy of this signal beyond the spatial frequency *B* is negligible³. We thus sample this signal on a discrete uniform grid of N = 2LB points and represent the resulting discrete signal by a vector $\rho \in \mathbb{C}^N$. The vector ρ is assumed to be *K*-sparse in an orthonormal basis $\Psi \in \mathbb{C}^{N \times N}$. We also consider a 1-D linear chirp, with chirp rate $w \in \mathbb{R}$, that reads as a complex-valued function $c(x) = e^{i\pi w x^2}$. On the field-of-view *L*, this linear chirp is approximately band limited at its maximum instantaneous frequency |w|L/2. This band limit can be parametrized in terms of a discrete chirp rate $\bar{w} = wL^2/N$ and thus $|w|L/2 = |\bar{w}|B$.

In this setting, the s_2MRI measurement model is given by (2) with

$$\Phi^{\star} = \mathsf{F}^{\star}\mathsf{C}\mathsf{U} \in \mathbb{C}^{N_{\mathrm{c}} \times N}.$$
(6)

In the above equation, the matrix U represents an up-sampling operator needed to avoid aliasing of the modulated signal due to a lack of sampling resolution in a digital description of the originally analog problem. Indeed, the convolution in Fourier space induced by the analog chirp modulation implies that the band limit of the modulated signal is the sum of the individual band limits of the original signal and of the chirp c. Therefore, an up-sampled grid with at least $N_c = (1 + |\bar{w}|)N$ points needs to be considered and the modulated signal is correctly obtained by applying the chirp modulation on the signal after up-sampling on the N_c points grid. The up-sampling operator U, implemented in Fourier space by zero padding, is thus of size $\mathbb{C}^{N_c \times N}$ and satisfies $U^*U = I \in \mathbb{C}^{N \times N}$. The matrix $C \in \mathbb{C}^{N_c \times N_c}$ is the diagonal matrix implementing the chirp modulation on the up-sampled grid and the matrix $F = \{f_i\}_{1 \le i \le N_c}$ stands for the discrete Fourier basis on the same grid.

The s₂MRI sensing matrix Φ is not orthogonal because of the presence of the matrix U. Consequently, the recovery condition (4) does not strictly hold. However, one can obtain a similar recovery condition [20]. In particular, if

$$M \ge DN_{\rm c} \ \mu_{\rm c}^2(\Phi^\star, \Psi) K \log^4(N) \tag{7}$$

for some universal constant D, then the vector $\boldsymbol{\rho}$ is accurately recovered with high probability by solving the BP problem (3). This result shows that the number of measurements M needed to reconstruct K-sparse signals is proportional to the product $N_c \ \mu_c^2(\Phi^*, \Psi)$ rather than to $\ \mu_c^2(\Phi^*, \Psi)$ only. The s₂MRI technique is efficient only if this product decreases with the chirp rate \bar{w} . In other words, the number of measurements needed for accurate recovery of sparse signals decreases only for sparsity basis Ψ for which the mutual coherence $\ \mu_c^2(\Phi^*, \Psi)$ decreases faster than N_c increases.

C. Illustration

We illustrate here the effect of the chirp modulation on the number of measurements needed for accurate recovery of K-sparse signals.

We consider a 1-D complex signal ρ of size N = 256 corresponding to one line of an MRI brain image. This signal is decomposed into three different sparsity bases Ψ and hard-thresholded at a fixed sparsity K = 25. The sparsity bases considered are the Dirac basis, the Haar wavelet basis and the Fourier basis. This signal is then probed according to relation (2) with $\Phi^* = F^*CU$ and $\bar{w} \in \{0, 0.1, 0.3, 0.5\}$, and reconstructed from different numbers of measurements M by solving the BP problem (3). Each time, the probability of recovery⁴ of the signal is computed over 1000 simulations. In this experiment, no noise is added to the measurements \boldsymbol{y} , and the indices i of the selected Fourier basis vectors \boldsymbol{f}_i are chosen uniformly at random from $\{0, \ldots, N_c - 1\}$.

The BP reconstructions in the presence of chirp modulation with chirp rate \bar{w} are denoted by BP \bar{w} . For each sparsity basis and chirp rate considered, the curves of the probability of recovery as a function of the number of measurements M are reported in Fig. 2. The corresponding values of the product $N_c \ \mu_c^2 (\Phi^*, \Psi)$ are reported in Table I.

For the Dirac basis, the number of measurements needed to reach a probability of recovery of 1 slightly increases with the chirp rate, as suggested by the values of $N_c \mu_c^2(\Phi^*, \Psi)$ in Table I and relation (7). On the contrary, for the Haar and Fourier bases, the values of $N_c \mu_c^2(\Phi^*, \Psi)$ in the table predict a drastic improvement of the results with an increase of the chirp rate. This prediction is confirmed by the results presented in Fig. 2. Note that for the Haar and Fourier bases, the value of the product

³In Section III, we will introduce a setting that properly takes into account the fact that, in general, MRI signals are not band limited.

⁴Perfect recovery is considered to occur if the ℓ_2 -norm between the original signal ρ and the reconstructed signal ρ^* satisfies: $\|\rho - \rho^*\|_2 \le 10^{-3} \|\rho\|_2$



Fig. 2. Probability of recovery of the signal ρ as a function of the number of measurements M for $\bar{w} = 0$ (dotted black curve), $\bar{w} = 0.1$ (continuous blue curve), $\bar{w} = 0.3$ (continuous red curve), $\bar{w} = 0.5$ (continuous black curve), and three sparsity bases: the Dirac basis (left); the Haar wavelet basis (middle); the Fourier basis (right).

TABLE I VALUES OF $N_{\rm c}~\mu_{\rm c}^2$ at Different Chirp Rates \bar{w}

$N_{ m c}\mu_{ m c}^2$	Dirac	Haar	Fourier
$\bar{w} = 0$	1.00	256	256
$\bar{w} = 0.1$	2.27	43.5	15.5
$\bar{w} = 0.3$	2.97	25.9	6.11
$\bar{w} = 0.5$	3.29	22.4	4.17

 $N_c \ \mu_c^2(\Phi^*, \Psi)$ decreases much less between $\overline{w} = 0.3$ and $\overline{w} = 0.5$ than between $\overline{w} = 0.1$ and $\overline{w} = 0.3$, suggesting a smaller improvement in the number of measurements needed for accurate recovery. This is also in line with the curves of the probability of recovery in Fig. 2. In summary, as predicted by the theory, the effect of the s₂MRI technique depends on the sparsity basis. Note that the decrease of the performance for the Dirac basis (optimally incoherent only at $\overline{w} = 0$) is negligible compared to the improvement obtained for the two other bases. Note also that MRI signals are usually sparse in wavelet bases [1]. The results obtained with the Haar wavelet basis therefore suggest strong efficiency of the technique in MRI.

One can wonder if the proposed encoding scheme can result in shift-variant reconstruction quality. Indeed, the phase variation at the center of the chirp is less rapid than at the limit of the field-of-view. Let us consider the case of a signal sparse in the Haar wavelet basis. This signal can, roughly, be separated in large scale and fine scale sparsity basis vectors. The region where the chirp is oscillating slowly is limited to a small part of the field-of-view. Consequently, large scale sparsity basis vectors are necessarily affected by the chirp modulation as they have a wide support in signal space. The energy of these basis vectors is thus spread in k-space thus improving the reconstruction quality of the low scale structures of the signal. On the other hand, the fine scale sparsity basis vectors at the center of the field-of-view remains not significantly modulated. However, these vectors are already incoherent with the Fourier basis as their energy naturally spreads out in k-space. The fine scale structure of the signal are well recovered in absence and presence of chirp modulation. In summary, almost no shift-variant reconstruction quality is to be expected.

Let us acknowledge that the idea of convolving the k-space to optimize the acquisition procedure in the context of compressed sensing can also be found in [16]–[18]. In these works, the k-space is convolved by a random Toeplitz matrices. We should also note that the spread spectrum technique can be related to the random convolution approach where the signal is convolved by a random sequence and under-sampled in *real* space [40]. In our case, convolution and under-sampling occur in *k*-space. Finally, let us mention that in a discrete setting, i.e., in the absence of band limit extension, replacing the linear chirp modulation by a random modulation leads to a universal encoding strategy, i.e., the reconstruction quality does not depend on the sparsity basis [20]. The s₂MRI technique tries to emulate this universal encoding strategy. A universal compressed sensing strategy might as well be obtained by projecting the signal ρ onto random waveforms with Gaussian distributions [34], [35]. In the context of MRI, an encoding strategy based on this sensing scheme was recently proposed in [19].

III. SPREAD SPECTRUM MRI (S₂MRI)

A. Quadratic Phase Profiles

The spread spectrum principle was explained in the previous section in a simplified analog framework. Here, we move to a setting encompassing realistic analog MR signals which are limited in field-of-view and are consequently *not* band limited.

Standard MR measurements take the form of k-space coefficients, with $k = (k_x, k_y, k_z)$, of the original three-dimensional image probed representing the tissue magnetization as a function of the position inside a given field-of-view. This image is complex-valued due to magnetic field inhomogeneities [41]. We consequently denote it by a complex-valued function $\rho(x)$ of the position $x \in \mathbb{R}^3$, with components (x, y, z) in image space. (x, y) conventionally corresponds to the phase encoding directions, and z corresponds to the readout direction. The field-of-view is $L = L_x \times L_y \times L_z$. We also consider quadratic phase profiles represented by a linear chirp in the phase encoding directions

$$c(\boldsymbol{x}) = \mathrm{e}^{\mathrm{i}\pi(w_x x^2 + w_y y^2)} \tag{8}$$

with chirp rate $(w_x, w_y) \in \mathbb{R}^2$. This linear chirp is characterized by an instantaneous frequency $(w_x x, w_y y)$ at position x. On the finite field-of-view L, it is therefore approximately a band-limited function with approximate band-limits $(|w_x|L_x/2, |w_y|L_y/2)$ in the phase encoding directions.

In this setting, MR measurements at spatial frequency $\mathbf{k} \in \mathbb{R}^3$ take the general form

$$\nu(\boldsymbol{k}) = \int_{\mathbb{R}^3} \rho(\boldsymbol{x}) c(\boldsymbol{x}) \mathrm{e}^{-2\mathrm{i}\pi \boldsymbol{k} \cdot \boldsymbol{x}} \mathrm{d}^3 \boldsymbol{x}.$$
 (9)

In other words, the measurement ν corresponds to the coefficient at spatial frequency \boldsymbol{k} of a signal obtained as the product of the original image $\rho(\boldsymbol{x})$ with the linear chirp modulation $c(\boldsymbol{x})$. In the absence of modulation $(w_x = w_y = 0)$, the measurements simply reduce to standard \boldsymbol{k} -space measurements. In the presence of quadratic phase profiles, the modulation amounts to the convolution of the Fourier transform of the chirp with that of the original image.

B. Under-Sampling in k-Space and the Inverse Problem

Let us assume that we want to probe the signal ρ at a resolution corresponding to a band-limit $B = (B_x, B_y, B_z)$. Note that it does not imply that the signal is band-limited and some energy may remain beyond B. At this resolution, the signal of interest is discretized on a discrete uniform grid of $N = N_x \times N_y \times N_z$ spatial frequencies $\mathbf{k}_i \in \mathbb{R}^3$, $1 \le i \le N$, in \mathbf{k} -space, with $N_x = 2L_x B_x$, $N_y = 2L_y B_y$, and $N_z = 2L_z B_z$. In real space, this discretized signal may equivalently be described by its coefficients on a discrete uniform grid of N points \mathbf{x}_i with $1 \le i \le N$. On this discrete grid, the linear chirp can be parametrized in terms of discrete chirp rates $(\bar{w}_x, \bar{w}_y) = (w_x L_x^2 / N_x, w_y L_y^2 / N_y)$, thus exhibiting approximate band-limits $(|\bar{w}_x|B_x, |\bar{w}_y|B_y, 0)$ on the finite field-of-view L.

We assume that the spatial frequencies probed span the k-space up to the band-limits B. We also assume that these frequencies belong to the discrete grid of points k_i , so that we can avoid any re-gridding operation [41]. Due to the linear chirp modulation, the measurements can contain significant energy from k-space coefficients beyond the band-limits B. Considering the estimated band-limits of the linear chirp, we choose to reconstruct the original signal ρ on a high resolution grid of $N_c = N_x(1 + |\bar{w}_x|) \times N_y(1 + |\bar{w}_y|) \times N_z$ points in order to prevent any aliasing in the reconstruction algorithm from these high spatial frequencies. The sampled signal on this high resolution grid is denoted by a vector $\rho \in \mathbb{C}^{N_c}$. The reconstructed signal is subsequently down-sampled at the desired resolution by keeping only its spatial frequencies belonging to the previously defined grid of size N.

The **k**-space coverage provided by the *M* spatial frequencies probed \mathbf{k}_b , with $1 \leq b \leq M$, can be represented by a binary mask in **k**-space, equal to 1 for each spatial frequency probed and 0 otherwise. The measurements may be denoted by a vector of *M* **k**-space coefficients $\mathbf{v} = \{v(\mathbf{k}_b)\}_{1 \leq b \leq M} \in \mathbb{C}^M$, possibly affected by noise, which is denoted by the vector $\mathbf{n} = \{n_b\}_{1 \leq b \leq M} \in \mathbb{C}^M$. In this setting, we consider an incomplete \mathbf{k} -space coverage (M < N), in order to accelerate the acquisition time in comparison with a complete \mathbf{k} -space coverage. In order to satisfy standard MRI constraints, an arbitrary under-sampling can be considered in the phase encoding directions while all spatial frequencies in the readout direction k_z are probed. This form of under-sampling can directly be expressed in terms of an acceleration of the acquisition. Thus, the s₂MRI measurement model satisfies

$$\boldsymbol{\nu} = \mathsf{MF}^{\star}\mathsf{CU}\boldsymbol{\rho} + \boldsymbol{n} \tag{10}$$

where the matrix $\mathsf{MF}^{\star}\mathsf{CU} \in \mathbb{C}^{M \times N_c}$ encodes the complete linear relation between the signal and the measurements. The rectangular matrix $\mathsf{U} \in \mathbb{C}^{N_u \times N_c}$ represents an up-sampling operation, implemented in \mathbf{k} -space space by zero padding. This operation is needed to avoid any aliasing of the modulated signal due to a lack of sampling resolution in a discrete description of the originally continuous problem. The modulated signal is correctly obtained by applying the chirp modulation to the signal after up-sampling on the grid of $N_u = N_x(1+2|\bar{w}_x|) \times N_y(1+$ $2|\bar{w}_y|) \times N_z$ points by zero padding in the k_x and k_y directions. The matrix $\mathsf{C} \in \mathbb{C}^{N_u \times N_u}$ is the diagonal matrix implementing the chirp modulation on the up-sampled grid. The unitary matrix $\mathsf{F} \in \mathbb{C}^{N_u \times N_u}$ stands for the discrete Fourier basis on this high resolution grid. The matrix $\mathsf{M} \in \mathbb{R}^{M \times N_u}$ is the rectangular binary matrix implementing the mask. It contains only one nonzero value on each line, at the index of the k-space coefficient corresponding to each of the spatial frequencies probed.

Regarding the reconstruction of the signal ρ , (10) represents the measurement constraint. We take a statistical point of view and consider independent Gaussian noise on each measurement. Considering a candidate reconstruction $\bar{\rho}$, the residual noise reads as $\bar{n} = \nu - MF^*CU\bar{\rho}$. The noise level estimator, defined as twice the negative logarithm of the likelihood associated with $\bar{\rho}$, reads as

$$\chi^{2}(\bar{\boldsymbol{\rho}}, \boldsymbol{\nu}) = \sum_{b=1}^{M} \frac{|\bar{n}_{b}|^{2}}{[\sigma^{(n_{b})}]^{2}}$$
(11)

where the symbol $|\cdot|$ stands for the complex norm and $\sigma^{(n_b)}$ for the standard deviation of each of the real and imaginary parts of the noise component n_b . This noise level estimator follows a chi-square distribution with 2M degrees-of-freedom. Typically, this estimator should be minimized by a good candidate for the reconstruction. As explained in [33], the measurement constraint on the reconstruction may be defined as a bound $\chi^2(\bar{\boldsymbol{p}}, \boldsymbol{\nu}) \leq \epsilon^2$, with ϵ^2 corresponding to some large percentile of the chi-square distribution. In the remainder of the paper, we choose the 99th percentile.

C. Variable Density Sampling

As discussed in [1], the method selecting the spatial frequencies \mathbf{k}_b , with $1 \leq b \leq M$, is critical to achieve good reconstruction. Their approach is based on the use of a variable density sampling method. The distributions, which are empirically identified to provide optimized reconstructions, are defined by a power law decaying function $f(\mathbf{k}) = (1 - |\mathbf{k}|/|\mathbf{k}_m|)^p + \beta$ for some power p > 0 and real constant $\beta \in [-1, 1]$ ($|\mathbf{k}|^2 = \mathbf{k}_x^2 + \mathbf{k}_y^2$ and \mathbf{k}_m is the highest spatial frequency of the \mathbf{k} -space domain to probe). In the present work, the spatial frequencies \mathbf{k}_b are selected using the method⁵ of [1]: the values $f(\mathbf{k}_i)$ are first thresholded to restrict them to the interval [0, 1]; N independent binary random variables $\delta(\mathbf{k}_i)$ taking value 1 with probability $f(\mathbf{k}_i)$ are then generated; and finally the mask M selecting the \mathbf{k} -space coefficients for which $\delta(\mathbf{k}_i) = 1$ is created.

Let us highlight the importance of the value β on the actual shape of the variable density profile. For a fixed power p, this constant is computed beforehand in order to ensure that the number of measurements is, on average, equal to its target value M. Given a fixed number M of measurements, we denote p_M the value of p for which $\beta = 0$. For a small value of $p(p < p_M)$, one would intuitively predict that the measurements spread all over the k-space domain with the density of points higher at the center of k-space. In fact, for $p < p_M$, one has $\beta < 0$ and a whole **k**-space region at high frequency remains unprobed. On the contrary, we have $\beta > 0$ for $p > p_M$ and the **k**-space is probed with a nonzero probability at the edges. Consequently, $p = p_M$ is the first power that ensures that the entire k-space domain is probed with a nonzero probability. Also note that for $p > p_M$, the center of the *k*-space is fully sampled, as $\beta > 0$, and that the size of the fully sampled region increases when p increases.

In practice, the choice of the power p of the variable density profile is difficult and should be adapted with the number

 s_2MRI



Variable density sampling

Fig. 3. Original brain image at 1 mm of resolution. The first panel represents its absolute value. The second panel represents its phase mapped between $-\pi$ and π . The last two panels represent the logarithm of the amplitude of its Fourier transform before (third panel) and after (fourth panel) chirp modulation with chirp rate $\bar{w} = 0.3$. Dark and light regions respectively indicate low and high intensities.

of measurements in order to obtain the best reconstruction qualities. In the absence of chirp modulation, we performed reconstructions for several values of p and noticed that $p = p_M$ leads to the best performance⁶. In the presence of a linear chirp modulation, the power spectrum of the original signal is spread and flattened. However, in the range of the chirp rates studied, the power spectrum of the modulated signal remains peaked at the origin (Fig. 3). We therefore also apply a variable density sampling and choose the power p to be p_M . Empirically, this value also leads to the best reconstruction qualities. Note that choosing $p = p_M$ seems reasonable, as, given the spread of the information in the phase encoding directions, one wants to distribute the measurements over the entire sampling region of size B and also limit the size of the fully sampled region at the **k**-space center.

Let us acknowledge that recent theoretical results obtained in [42] support the choice of such profiles in MRI.

D. Numerical Simulations

1) Simulation Protocol: To test the proposed technique, a real brain image is acquired on a 7T short bore, actively shielded, MR scanner (Siemens, Erlangen, Germany). The

Fig. 4. Simulated reconstructions for an under-sampling of 20% of N and an input snr of 32. The top row shows the reconstructions for the variable density sampling (left panel) and the s₂MRI (right panel) techniques, respectively. The bottom row shows the error images (difference between the original image and the reconstructions) in the absence (left panel) and presence (right panel) of chirp modulation. For a better visualization, the error images were scaled by a factor of 6. The colormap for the error images goes from white to black, indicating low and high errors, respectively.

subjects provided written informed consent prior to the imaging session, according to the guidelines of the local ethics committee. The parameters of the acquisition are as follows: $L_x \times L_y \times L_z = 224 \times 168 \times 208 \text{ mm}^3$ with a resolution of $0.5 \times 0.5 \times 4 \text{ mm}^3$. The matrix size is thus $N_x \times N_y \times N_z = 448 \times 336 \times 52$. The standard clinical MPRAGE sequence is used with echo time TE = 3.53 ms, inversion time TI = 1.3 s, repetition time TR = 3.5 s, and bandwidth BW = 200 Hz. Note that, for the sake of simplicity, we restrict our analyses to one two-dimensional z-slice of the original three-dimensional acquisition with an under-sampling in both phase encoding directions k_x and k_y . Also note that the image used is complex-valued.

In order to model an analog acquisition scheme, the original image at a resolution of 0.5 mm is used to compute the measurements but the reconstruction is performed at a resolution of 1 mm. The reconstructed images are compared to the image obtained with a full acquisition at 1 mm of resolution (Fig. 3).

The parameters of our analyses are as follows. Firstly, acquisitions are considered for various numbers M of complex measurements corresponding to coverages of 5%, 10%, 15%, 20%,

⁶In practice, p_M is determined by an iterative process. Starting from p = 0, we increase its value by 0.5 until we obtain $\beta \ge 0$ for the number of measurements considered.



Fig. 5. Relative reconstruction errors as functions of the input snr for the variable density sampling technique (dashed blue curve) and the s_2 MRI technique with $\bar{w} = 0.3$ (dot-dashed red curve) and $\bar{w} = \bar{w}_{opt}$ (continuous black curve). The first to sixth panels show the curves for coverages of 5%, 10%, 15%, 20%, 25%, and 50% of N, respectively. All curves represent the mean relative error over 30 simulations, and the vertical lines represent the error at 1 standard deviation.



Fig. 6. Relative reconstruction errors as functions of the input snr for the variable density sampling technique (dashed blue curve) and the s_2 MRI technique (dot-dashed red curve) with varying chirp modulation. From left to right and top to bottom, the panels show the curves for coverages of 15%, 25%, 50% of N respectively. All curves represent the mean relative error over five simulations, and the vertical lines represent the error at 1 standard deviation.

25%, and 50% of N. Secondly, instrumental noise is also added to the measurements as independent and identically distributed zero-mean Gaussian noise. The corresponding standard deviation σ is identical for all the frequencies probed and we consider values of input snr⁷ of 2^j , with $0 \le j \le 5$. Thirdly, the chirp modulation studied has the same chirp rate \bar{w} in both phase encoding directions, i.e., $\bar{w} = \bar{w}_x = \bar{w}_y$, with values in the range [0, 0.5]. Fourthly, the signals are reconstructed by solving the BP problem where the Total Variation (TV) norm of the signal ρ is substituted for the ℓ_1 norm⁸. This problem is solved thanks the Douglas Rachford algorithm [44], [45]. Note that the TV norm in combination with wavelet sparsity basis is, for example, used for reconstructing MRI signals from under-sampled k-space in [1], [6], or [14]. For our problem, we tested several multiscale representations such as Daubechies wavelets, steerable wavelets, or curvelets, but the best reconstructions were obtained with the TV norm. Finally, for each value of M and input snr considered, 30 simulations are generated with independent noise and mask realizations, and the relative reconstruction errors $\|\boldsymbol{\rho} - \boldsymbol{\rho}^{\star}\|_2 / \|\boldsymbol{\rho}\|_2$ are computed for the variable density sampling and s_2 MRI techniques. For each value of M and input snr, the discrete chirp rate $\bar{w}_{\rm opt}$ that gives the smallest relative error on average over the 30 simulations is recorded.9

Variable density sampling

 s_2MRI



Fig. 7. Simulated reconstructions with varying chirp modulation for an undersampling of 50% of N and an input snr of 32. The top row shows the reconstructions of an axial slice for the variable density sampling (left panel) and the s_2 MRI (right panel) techniques, respectively. The bottom row shows the error images (difference between the original image and the reconstructions) in the absence (left panel) and presence (right panel) of chirp modulation. For a better visualization, the error images were scaled by a factor of 8. The colormap for the error images goes from white to black, indicating low and high errors, respectively.

2) Simulation Results: The magnitudes of the reconstructed images obtained with the s_2MRI and the variable density sampling techniques for an acceleration factor of 5 and an input snr of 32 are presented in Fig. 4, along with the corresponding error images (magnitudes of the complex-valued differences between the original image and reconstructed images). The relative errors of the reconstructions as functions of the input snr for the

⁷The snr is defined as the ratio between the mean value of the complex magnitude of the original signal and the standard deviation of the noise σ .

⁸The TV norm of a signal is defined as the ℓ_1 norm of the magnitude of its gradient [32], [43]. Note that the recovery condition (7) does not hold with this norm. However, one can notice that the TV norm of a signal ρ is very similar to the ℓ_1 norm of its decomposition in the Haar wavelet basis. In the light of the preliminary results of Section II-C, one can thus hope to obtain an improvement of the reconstruction quality in presence of chirp modulation.

⁹s2MRI toolbox available at http://lts2www.epfl.ch/people/gilles.

six coverages considered and for both methods are reported in Fig. 5.

For acceleration factors larger than 2, the s₂MRI technique with $\bar{w} = 0.3$ provides better reconstruction than the variable density sampling technique with an improvement up to 0.05 of the relative error. Indeed, the relative error is, on average, lower in the presence of the chirp modulation. The corresponding standard deviations are also much smaller, indicating that the s₂MRI technique is more stable¹⁰. At an acceleration factor of 2, the variable density sampling technique gives slightly better reconstructions than the s₂MRI technique with $\bar{w} = 0.3$. However, with $\bar{w} = \bar{w}_{opt} = 0.1$, the s₂MRI technique provides relative errors similar to those obtained with the variable density sampling technique. These results suggest to reduce the chirp rate as the number of measurements increases. This is coherent with the fact that modulation is not needed in the limit of no under-sampling.

When comparing the magnitudes of reconstructed images in Fig. 4, differences between both methods do not appear obvious. However, one can notice that fine details are recovered better with the s_2MRI technique: the vessel (white spot) indicated by a blue arrow appears in the s_2MRI reconstruction, but not in the variable density sampling reconstruction. The error images bring more information, and one can notice that the errors are smaller in the presence of chirp modulation. In particular, the low scale structures, rendered incoherent with the chirp modulation, are better recovered.

IV. S₂MRI IMPLEMENTATION AND EXPERIMENTAL RESULTS

A. Implementation

The s_2 MRI technique is tested on the 7T scanner described in Section III-D with 3-D acquisitions of a phantom and a human brain. For the brain experiment, the subjects provided written informed consent prior to the imaging session, according to the guidelines of the local ethics committee.

The chirp modulation is implemented through the use of a second order shim coil $x^2 - y^2$. In our implementation, the chirp rate varies linearly with the readout time t (or equivalently k_z) and is proportional to the intensity of the quadratic magnetic field κ : $w(t) = w_x(t) = -w_y(t) = \gamma \kappa t/\pi$, where γ is the gyromagnetic factor. The maximum chirp rate w^{max} is reached at $t = \text{TE} + \Delta t/2$, the mean chirp rate w^{mean} at t = TE, and the minimum chirp rate w^{min} at $t = \text{TE} - \Delta t/2$ (Δt is the readout duration). This chirp modulation can be introduced in the measurement model by modifying (10) as follows:

$$\boldsymbol{\nu} = \mathsf{MF}_{xy}^{\star}\mathsf{CF}_{z}^{\star}\mathsf{U}\boldsymbol{\rho} + \boldsymbol{n}. \tag{12}$$

In the above equation, $\mathsf{F}_z^{\star} \in \mathbb{C}^{N_{\mathrm{u}} \times N_{\mathrm{u}}}$ and $\mathsf{F}_{xy}^{\star} \in \mathbb{C}^{N_{\mathrm{u}} \times N_{\mathrm{u}}}$ implement the Fourier transform along the *z*-direction and the *xy*-directions, respectively. The matrix $\mathsf{C} \in \mathbb{C}^{N_{\mathrm{u}} \times N_{\mathrm{u}}}$ implements the chirp modulation. Signals are thus reconstructed on a grid of $N_c = N_x(1 + |\bar{w}_x^{\max}|) \times N_y(1 + |\bar{w}_y^{\max}|) \times N_z$ points, and $N_u = N_x(1 + 2|\bar{w}_x^{\max}|) \times N_y(1 + 2|\bar{w}_y^{\max}|) \times N_z$.

This modulation can be decomposed as a quadratic phase modulation in the (x, y) planes with chirp rate w^{mean} , combined with a linear phase modulation in k_z . This linear phase modulation produces shifts of the original signal by an amount proportional to $\kappa(x^2 - y^2)$ along the z direction. The chirp modulation used is thus not ideal, as it creates distortions of the original object and complicates the measurement matrix. However, the energy of sparsity basis vectors is still spread by the main chirp modulation with chirp rate w^{mean} , and the previous conclusions based on theory and simulations should still hold. This will be confirmed by the numerical experiments of Section IV-B. Note also that the reconstructed images are free of any distortion as the complete effect of the modulation is modeled in (12). Let us remark that these distortions might be avoided with the use of RF pulses or dedicated coils applied only during phase encoding.

As in Section III-D, the s₂MRI technique is compared to the variable density sampling method with $p = p_M$. Full acquisitions (M/N = 1) are performed both in the absence and in the presence of the chirp modulation. The number of phase encodings (k_x, k_y) is then reduced retrospectively by applying a mask on the complete data.

The noise level is evaluated directly on the under-sampled data available for reconstruction. For each (k_x, k_y) pair measured, all the frequencies k_z are probed so that the signal is available as a function of z. The level of the noise is estimated on probed pairs (k_x, k_y) at positions z that do not contain any signal. This noise level is identical in the presence and absence of chirp modulation.

B. Numerical Validation

In this section, we perform simulations using the acquisition scheme described above to confirm that, even though the implementation of the chirp modulation is not ideal, the reconstruction quality is still enhanced with the s₂MRI technique.

For this numerical experiment, a brain volume was acquired using the standard clinical MPRAGE sequence on a field-ofview of $L_x = 243$ mm, $L_y = 176$ mm, and $L_z = 256$ mm, with a resolution of 1 mm in each direction ($N_x = 243$, $N_y = 176$, $N_z = 256$). The echo time is TE = 4.59 ms, the inversion time TI = 1.5 s, the repetition time TR = 3.5 s, the bandwidth BW = 250 Hz. As in Section III-D, in order to model an analog acquisition scheme, the original image at a resolution of 1 mm is used to compute the measurements, but the reconstruction is performed at a resolution of 2 mm. The reconstructed images are compared to the image obtained with a full acquisition at 2 mm of resolution.

The parameters of our experiment are as follows. Firstly, acquisitions are considered for various numbers M of complex measurements corresponding to coverages of 15%, 25%, 50% of N. Secondly, instrumental noise is also added to the measurements as independent and identically distributed zero-mean Gaussian noise. The corresponding standard deviation σ is identical for all the frequencies probed and we consider values of input snr of 2^j , with $0 \le j \le 5$. Thirdly, the simulated chirp

¹⁰Some leftover variability for the variable density sampling technique might still be removed by increasing the number of simulations. However, this would not modify the results and comparison with s_2 MRI.



Fig. 8. Phantom and brain reconstructions from real experimental data with the s_2 MRI technique. The first to fourth columns show the magnitude of the reconstructions from 15%, 25%, 50%, and 100% of phase encodings, respectively. The fifth column shows the reference images obtained by inverse Fourier transform (F.T.) of the fully sampled *k*-space.

modulation has a chirp rate varying linearly with k_z : $|\bar{w}_x| \in [0.12, 0.30]$ and $|\bar{w}_y| \in [0.09, 0.22]$. These values for the discrete chirp rate correspond to those used during the real experiment performed hereafter at 1 mm of resolution. However, relatively to the band-limit at 2 mm of resolution, the spectrum is naturally more spread than relatively to the band-limit at 1 mm of resolution. Therefore, for a reconstruction at 2 mm of resolution, the spectrum does not need to be spread as much as for a reconstruction at 1 mm of resolution. We thus divided the

values of the chirp rate by 2 in both dimensions. Fourthly, the signals are reconstructed by solving the BP problem where the TV norm of the signal ρ is substituted for the ℓ_1 norm. Finally, for each value of M and input snr considered, five simulations are generated with independent noise and mask realizations, and the relative reconstruction errors are computed for the variable density sampling and s₂MRI techniques.

The relative errors of the reconstructions as functions of the input snr for the three coverages considered and for both



Fig. 9. Phantom and brain reconstructions from real experimental data with the s_2 MRI technique (first and third rows, respectively) and the variable density sampling (VDS) technique (second and fourth rows, respectively). The first to fourth columns show the magnitude of the reconstructions from 15%, 25%, 50%, and 100% of phase encodings respectively. The fifth column shows the reference images obtained by inverse Fourier transform (F.T.) of the fully sampled *k*-space. The white lines on these images indicate the location of the spatial profiles presented in Fig. 10.

methods are reported in Fig. 6. The magnitudes of a reconstructed axial slice obtained with the s_2MRI and the variable density sampling techniques for an acceleration factor of 5 and an input snr of 32 are presented in Fig. 7, along with the corresponding error images. Conclusions of Section III-D still hold with this acquisition scheme. The relative error of reconstruction is lower in the presence of the chirp for the three coverage considered. When comparing the error images in Fig. 7, one can once more notice that the errors are smaller with the s_2MRI technique.

C. Experiments

For the phantom experiment, a gradient echo sequence is used on a field-of-view of $L_x = L_y = L_z = 192$ mm in the three directions with a resolution of 1 mm ($N_x = N_y = N_z = 192$). The echo time is TE = 6 ms, the repetition time TR = 10 ms, the bandwidth BW = 400 Hz, and the quadratic magnetic field intensity $\kappa = 3000 \ \mu \text{T/m}^2$. The discrete chirp rates satisfy $|\bar{w}_x| = |\bar{w}_y| \in [0.23, 0.36]$.

For the brain experiment, the standard clinical MPRAGE sequence is used on a field-of-view of $L_x = 243 \text{ mm}$, $L_y = 176 \text{ mm}$ and $L_z = 256 \text{ mm}$, with a resolution of 1 mm in each direction ($N_x = 243$, $N_y = 176$, $N_z = 256$). The echo time is TE = 4.59 ms, the inversion time TI = 1.5 s, the repetition time TR = 3.5 s, the bandwidth BW = 250 Hz, and the quadratic magnetic field intensity $\kappa = 4500 \mu \text{T/m}^2$. The discrete chirp rates satisfy $|\bar{w}_x| \in [0.25, 0.61]$ and $|\bar{w}_y| \in [0.18, 0.44]$.

Slices of the 3-D reconstructions obtained with the s_2MRI technique for 15%, 25%, 50%, and 100% of phase encodings are presented in Fig. 8 for the phantom and brain acquisitions.



Fig. 10. Phantom and brain reconstruction profiles from real experimental data with the s_2 MRI technique (first and third rows, respectively) and the variable density sampling (VDS) technique (second and fourth rows, respectively). The first to fourth columns show the profile magnitude of the reconstructions (continuous red curve) from 15%, 25%, 50%, and 100% of phase encodings respectively as well as the reference profiles (dotted black curve) obtained by inverse Fourier transform (F.T.) of the fully sampled **k**-space.

The images obtained by inverse Fourier transform of the fully sampled k-space (after correction of the distortions due to the small shifts in the z direction) are also presented. Fig. 9 provides a comparison of the reconstructed images obtained with the s₂MRI and the variable density sampling techniques. In the aim of providing further insight on how each method preserves image resolution or, in other words, captures shape and magnitude of high frequency features, we also provide 1-D spatial profiles in Fig. 10. The white lines in Fig. 9 indicate the location of these profiles.

Firstly, as one would expect, the visual reconstruction quality improves when the number of phase encodings M increases.

In the limit of a coverage of 100%, we cannot identify any loss of details between the reconstructed images and the ones obtained by inverse Fourier transform. Moreover, the reconstructed image contains much less noise than the image obtained by inverse Fourier transform. Indeed, at a coverage of 100%, the problem (3) is essentially reduced to a denoising problem.

Secondly, as in Section III-D, the differences between both methods do not appear at first glance when comparing only the magnitudes of reconstructed images in Fig. 9. Unfortunately here, the ground truth image is not accessible, so the corresponding error images cannot be displayed. However, a thorough visual inspection reveals that, for acceleration factors

Enhanced resolution

Standard resolution



Fig. 11. Axial (x, y) reconstructed slice for a coverage of 100% in the presence of chirp modulation on a high resolution grid of size $N_c = 393 \times 255$ (right panel) and after dowsampling on the original grid of size $N = 243 \times 176$.

larger than 2, some fine details are better recovered by our approach. On the phantom reconstructions for acceleration factors of 6.7 and 4 (coverages of 15% and 25%, respectively), the separation between the two biggest circles is more visible. The shapes of the circles are also more curved. On the brain reconstructions for an acceleration factor of 6.7, the vessels at the center of image are still visible with our method but not with the variable density sampling technique. The cerebral cortex also appears sharper. For an acceleration factor of 2, the thin layer separating the two hemispheres of the brain remains more visible with the s₂MRI technique.

Thirdly, the reconstructed spatial profiles of the phantom presented in Fig. 10 show that the shape and magnitude of high frequency features are, more often, slightly better recovered with the s_2MRI technique for acceleration factors larger than 4 (blue and red arrows indicate features better reconstructed with s_2MRI and variable density sampling respectively). This improvement is much more significant on the brain data (see arrows), and holds for acceleration factors larger than 2.

Finally, the distortions are correctly taken into account by the operator C, as both fully sampled slices of the phantom with and without chirp are identical (see Fig. 9). For the brain images, some differences remain due to small movements of the subject between both acquisitions. Let us remark that no fitting of the chirp parameters (center and rates) was performed to improve on the theoretical values. This highlights sufficient stability of the technique relative to parameter approximations.

V. CONCLUSION AND DISCUSSION

We presented a spread spectrum technique (s_2MRI) designed to accelerate MR acquisitions by compressed sensing. It consists of premodulating the original image by a linear chirp, which results from the application of quadratic phase profiles, and then performing random *k*-space under-sampling. Nonlinear algorithms promoting signal sparsity are then used for image reconstruction.

In the context of compressed sensing theory, the effectiveness of the technique is supported by a decrease of coherence between the sensing and sparsity bases due to the premodulation. Simulations in a simplified analog setting confirm that the enhancement of the image reconstruction quality is linked to the evolution of the mutual coherence. The s₂MRI technique was compared with the state-of-the-art variable density sampling using realistic numerical simulations and real acquisitions. Simulation results shows that the s₂MRI technique performs slightly better than the variable density sampling technique in terms of relative reconstruction error for acceleration factors larger than 2. The chirp modulation was also implemented on a 7T scanner with the use of a second order shim coil. Simulations of this implementation confirms again the slight superiority of s₂MRI. Visual inspection of reconstructions obtained from real acquisition of phantom and *in vivo* data also shows that this first (nonideal) implementation provides slightly better reconstruction qualities than the variable density sampling method. The s₂MRI technique thus outperforms the variable density sampling technique in terms of all the criteria used for evaluation.

Regarding future evolutions of the s_2MRI technique, an implementation of the linear chirp modulation with the use of RF pulses or dedicated coils could simplify the measurement scheme by applying the chirp modulation only during phase encoding in order to avoid object distortions, and, in turn, further enhance the reconstruction quality. Moreover, fitting the effective chirp center and rates on the basis of the data could improve the measurement model and result in better reconstructions.

Let us also emphasize the potential interest of the s₂MRI technique from an enhanced resolution perspective as, in the presence of the chirp modulation, the original image is reconstructed at a high resolution in order to avoid any aliasing problems. One can indeed consider reaching a higher spatial resolution for a fixed acquisition time without probing higher spatial frequencies in practice, which would require stronger gradient coils. In this context, any regularization approach adding a sparsity prior can help to synthesize spatial frequency information higher than that contained in the data. But the chirp modulation implies that high spatial frequency information is actually probed at lower frequencies. For illustration, Fig. 11 shows a reconstructed (x, y) slice for a coverage of 100% on a high resolution grid $N_{\rm c} = 393 \times 255$ as well as the image obtained after downsampling on the grid $N = 243 \times 176$. One can notice that the high resolution image provides sharper details with fewer aliasing artifacts. In particular, the vessels are better resolved.

ACKNOWLEDGMENT

The authors would like to thank D. Shuman for his helpful suggestions to improve the writing quality of the paper, as well as the reviewers for their very useful comments.

REFERENCES

- M. Lustig, D. Donoho, and J. M. Pauly, "Sparse MRI: The application of compressed sensing for rapid MR imaging," *Magn. Reson. Med.*, vol. 58, pp. 1182–1195, 2007.
- [2] H. Jung, J. C. Ye, and E. Y. Kim, "Improved k-t BLAST and k-t SENSE using FOCUSS," *Phys. Med. Biol.*, vol. 52, pp. 3201–3226, 2007.
- [3] U. Gamper, P. Boesiger, and S. Kozerke, "Compressed sensing in dynamic MRI," Magn. Reson. Med., vol. 59, pp. 365–373, 2008.

- [4] H. Jung, K. Sung, K. Nayak, E. Y. Kim, and J. C. Ye, "k-t FOCUSS: A general compressed sensing framework for high resolution dynamic MRI," *Magn. Reson. Med.*, vol. 61, pp. 103–116, 2009.
- [5] M. Usman, C. Prieto, T. Schaeffter, and P. G. Batchelor, "k-t Group Sparse: A Method for Accelerating Dynamic MRI," *Mang. Reson. Med.*, vol. 66, pp. 1163–1176, 2011.
- [6] D. Liang, B. Liu, J. Wang, and L. Ying, "Accelerating sense using compressed sensing," *Magn. Reson. Med.*, vol. 62, pp. 1574–1584, 2009.
- [7] M. Lustig and J. Pauly, "Spirit: Iterative self-consistent parallel imaging reconstruction from arbitrary k-space," *Magn. Reson. Med.*, vol. 64, pp. 457–471, 2010.
- [8] R. Otazo, D. Kim, L. Axel, and D. Sodickson, "Combination of compressed sensing and parallel imaging for highly accelerated first-pass cardiac perfusion MRI," *Magn. Reson. Med.*, vol. 64, pp. 767–776, 2010.
- [9] S. Hu, M. Lustig, A. P. Chen, J. Crane, A. Kerr, D. A. Kelley, R. Hurd, J. Kurhanewicz, S. J. Nelson, J. M. Pauly, and D. B. Vigneron, "Compressed sensing for resolution enhancement of hyperpolarized ¹³C flyback 3-D-MRSI," *Magn. Reson. Med.*, vol. 192, pp. 258–264, 2008.
- [10] S. Hu, M. Lustig, A. P. Chen, A. Balakrishnan, P. E. Z. Larson, R. Bok, J. Kurhanewicz, S. J. Nelson, A. Goga, J. M. Pauly, and D. B. Vigneron, "3-D compressed sensing for highly accelerated hyperpolarized ¹³C MRSI with in vivo applications to transgenic mouse models of cancer," *Magn. Reson. Med.*, vol. 63, pp. 312–321, 2010.
- [11] P. E. Z. Larson, S. Hu, M. Lustig, A. B. Kerr, S. J. Nelson, J. Kurhanewicz, J. M. Pauly, and D. B. Vigneron, "Fast dynamic 3-D MR spectroscopic imaging with compressed sensing and multiband excitation pulses for hyperpolarized (13)C studies," *Magn. Reson. Med.*, vol. 65, pp. 610–619, 2010.
- [12] S.-J. Kim, K. Koh, M. Lustig, S. Boyd, and D. Gorinevsky, "An interior-point method for large-scale ℓ₁-regularized least squares," *IEEE J. Sel. Topics Signal Process.*, vol. 1, pp. 606–617, 2007.
- [13] J. Trzasko and A. Manduca, "Highly undersampled magnetic resonance image reconstruction via homotopic ℓ₀-minimization," *IEEE Trans. Med. Imag.*, vol. 28, no. 1, pp. 106–121, Jan. 2009.
- [14] M. Guerquin-Kern, M. Häberlin, K. P. Pruessmann, and M. Unser, "A fast wavelet-based reconstruction method for magnetic resonance imagings," *IEEE Trans. Med. Imag.*, vol. 30, no. 9, pp. 1649–1660, Sep. 2011.
- [15] M. Seeger, H. Nickisch, R. Pohmann, and B. Scholkopf, "Optimization of k-space trajectories for compressed sensing by Bayesian experimental design," *Magn. Reson. Med.*, vol. 63, no. 1, pp. 116–126, 2010.
- [16] F. Sebert, Y. M. Zou, and L. Ying, "Compressed sensing MRI with random B1 field," in *Proc. ISMRM*, 2008, p. 3151.
- [17] D. Liang, G. Xu, H. Wang, K. F. King, D. Xu, and L. Ying, "Toeplitz random encoding MR imaging using compressed sensing," in *Proc. IEEE Int. Symp. Biomed. Imag. Nano to Macro (ISBI)*, 2009, pp. 270–273.
- [18] H. Wang, D. Liang, K. King, and L. Ying, "Toeplitz random encoding for reduced acquisition using compressed sensing," in *Proc. ISMRM*, 2009, p. 2669.
- [19] J. Haldar, D. Hernando, and Z.-P. Liang, "Compressed-sensing in MRI with random encoding," *IEEE Trans. Med. Imag.*, vol. 30, no. 4, pp. 893–903, Apr. 2011.
- [20] G. Puy, P. Vandergheynst, R. Gribonval, and Y. Wiaux, "Universal and efficient compressed sensing strategy through spread spectrum modulation," *EURASIP J. Adv. Signal Process.*, 2011.
- [21] Y. Wiaux, G. Puy, Y. Boursier, and P. Vandergheynst, "Spread spectrum for imaging techniques in radio interferometry," *Mon. Not. R. Astron. Soc.*, vol. 400, pp. 1029–1038, 2009.
- [22] A. A. Maudsley, "Dynamic range improvement in NMR imaging using phase scrambling," J. Magn. Reson., vol. 76, pp. 287–305, 1988.
- [23] V. J. Wedeen, Y. S. Chao, and J. L. Ackerman, "Dynamic range compression in MRI by means of a nonlinear gradient pulse," *Magn. Reson. Med.*, vol. 6, pp. 287–295, 1988.

- [24] J. G. Pipe, "Spatial encoding and reconstruction in MRI with quadratic phase profiles," *Magn. Reson. Med.*, vol. 33, pp. 24–33, 1995.
- [25] S. Ito and Y. Yamada, "Alias-free image reconstruction using fresnel transform in the phase-scrambling Fourier imaging technique," *Magn. Reson. Med.*, vol. 60, pp. 422–430, 2008.
- [26] F. M. Naini, R. Gribonval, L. Jacques, and P. Vandergheynst, "Compressive sampling of pulse trains: Spread the spectrum!," in *Proc. IEEE Int. Conf. Acoustics, Speech Signal Process.*, Taipei, 2009, pp. 2877–2880.
- [27] Y. Wiaux, G. Puy, Y. Boursier, and P. Vandergheynst, "Compressed sensing for radio interferometry: Spread spectrum imaging techniques," in *Proc. SPIE Conf. Wavelet XIII*, 2009, vol. 7446, p. 74460J.
- [28] J. D. McEwen and Y. Wiaux, "Compressed sensing for wide-field radio interferometric imaging," *Mon. Not. R. Astron. Soc.*, vol. 413, pp. 1318–1332, 2011.
- [29] G. Puy, Y. Wiaux, R. Gruetter, J.-P. Thiran, D. Van de Ville, and P. Vandergheynst, "Spread spectrum for interferometric and magnetic resonance imaging," in *Proc. IEEE Int. Conf. Acoustic, Speech Signal Process.*, 2010, pp. 2802–2805.
- [30] Y. Wiaux, G. Puy, R. Gruetter, J.-P. Thiran, D. Van de Ville, and P. Vandergheynst, "Spread spectrum for compressed sensing techniques in magnetic resonance imaging," in *Proc. IEEE Int. Sym. Biomed. Imag.*, 2010, pp. 756–759.
- [31] G. Puy, J. Marques, R. Gruetter, J.-P. Thiran, D. Van de Ville, P. Vandergheynst, and Y. Wiaux, "Accelerated MR imaging with spread spectrum encoding," in *Proc. ISMRM*, 2011, p. 2808.
- [32] E. J. Candès, J. Romberg, and T. Tao, "Robust uncertainty principles: Exact signal reconstruction from highly incomplete frequency information," *IEEE Trans. Inf. Theory*, vol. 52, no. 2, pp. 489–509, Feb. 2006.
- [33] E. J. Candès, J. Romberg, and T. Tao, "Stable signal recovery from incomplete and inaccurate measurements," *Comm. Pure Appl. Math.*, vol. 59, pp. 1207–1223, 2006.
- [34] E. J. Candès, "Compressive sampling," in Proc. Int. Congress Math., Madrid, Spain, 2006, vol. 3, pp. 1433–1452.
- [35] D. L. Donoho, "Compressed sensing," *IEEE Trans. Inf. Theory*, vol. 52, pp. 1289–1306, 2006.
- [36] E. J. Candès and J. Romberg, "Sparsity and incoherence in compressive sampling," *Inverse Problems*, vol. 23, pp. 969–985, 2007.
- [37] R. Baraniuk, "Compressive sensing," *IEEE Signal Process. Mag.*, vol. 24, pp. 118–12, 2007.
- [38] D. L. Donoho and J. Tanner, "Counting faces of randomly-projected polytopes when the projection radically lowers dimension," J. Amer. Math. Soc., vol. 22, pp. 1–53, 2009.
- [39] H. Rauhut, "Compressive sensing and structured random matrices," *Radon Series Comp. Appl. Math.*, vol. 9, pp. 1–92, 2010.
- [40] J. Romberg, "Compressive sensing by random convolution," SIAM J. Imag. Sci., vol. 02, pp. 1098–1128, 2009.
- [41] E. M. Haake, R. W. Brown, M. R. Thompson, and R. Venkatesan, Magnetic Resonance Imaging: Physical Principles and Sequence Design. New York: Wiley, 1999.
- [42] G. Puy, P. Vandergheynst, and Y. Wiaux, "On variable density compressive sampling," *IEEE Signal Process. Lett.*, vol. 18, no. 10, pp. 595–598, Oct. 2011.
- [43] L. I. Rudin, S. Osher, and E. Fatemi, "Nonlinear total variation based noise removal algorithms," *Physica D*, vol. 60, pp. 259–268, 1992.
- [44] H. H. Bauschke, R. S. Burachik, P. L. Combettes, V. Elser, D. R. Luke, and H. Wolkowicz, "Proximal splitting methods in signal processing," in *Fixed-Point Algorithms for Inverse Problems in Science and Engineering*. New York: Springer-Verlag, 2011.
- [45] M. J. Fadili and J.-L. Starck, "Monotone operator splitting for optimization problems in sparse recovery," in *Proc. ICIP*, Cairo, Egypt, 2009, pp. 1461–1464.